

Example ICPMS Reports

Result Report – Form 1
Initial/Continuing Calibration Verification Report—Form 2
Initial/Continuing Calibration Blank Report—Form 3A
Method Blank Report-Form 3B
Interference Check Standard A and AB Report—Form 4
Matrix Spike/Matrix Spike Duplicate Recovery Report-Form 5A
Post Spike Recovery Report-Form 5B
Duplicate RPD Report-Form 6
Laboratory Control Sample Recovery Report-Form 7
Serial Dilution % Difference Report-Form 8
Internal Standard Relative Intensity Summary Report—Form 15
Exception Report
Cover Page

Others not shown:

MS/MSD Recovery/RPD Report
LCS/LCSD Recovery/RPD Report
Reporting Limit Check Report
Linear Range Report
Detection Limit Report
Preparation Log
Analytical Sequence Run Logs
Internal Standard Associations
Tune Report
Initial Calibration Summaries
Analytical Checklist Report w/Electronic Signature
Notification Reports
Detection Summary Report

Note: The example is a generalized report set. The analytical reports may vary for laboratory preference, method requirements, regulatory requirements, and industry of need.

The example reports are used in EISC's:

- Express Report System
- Automated Quality Assurance System
- R & R Suite System (Level 4 deliverable system for potential liability and litigation type deliverables-Assembles EISC reports with instrument chromatograms).

Additionally, all report results can be transferred electronically (non-PDF and/or PDF) to information management systems, central repositories (Big Data Hubs), Excel ...





INORGANIC ANALYSIS DATA PACKAGE

Client: EISC Laboratories, Inc. **SDG No.:** **Deliverable ID** **Method Type:** ICPMS Method

Sample ID: Sample 1

Client ID: Client Sample 1

Matrix: SOIL **Date Received:** 10/15/2015 **Date Collected:** 10/14/2015 **Level:** LOW

% Solids: 89.2 **Sample Wt/Vol:** 1.0 **Final Vol:** 500.0

Prep Batch ID: 87566 **Prep Date:** 11/2/2015

Analyte	Concentration	Units	C	Qual	MDL	CRQL	Dil	Analytical		Instrument	Run	M
								Date	Time			
Antimony	<	0.23 mg/kg	U	*	0.23	1.1	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Arsenic		7.1 mg/kg		*	0.20	0.53	1	11/4/2015	14:30:14	ICPMS2	ICPMS Seq	MS
Barium		48.8 mg/kg		*	1.7	5.3	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Beryllium		0.69 mg/kg			0.17	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Cadmium		0.12 mg/kg	J		0.10	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Chromium		21.0 mg/kg		*	0.22	1.1	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Cobalt		6.7 mg/kg		*	0.12	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Copper		8.9 mg/kg			0.27	1.1	1	11/3/2015	15:19:37	ICPMS2	ICPMS Seq	MS
Lead		26.5 mg/kg			0.14	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Manganese		327 mg/kg		*	0.17	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Nickel		10.0 mg/kg		*	0.13	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Selenium		1.3 mg/kg	J		0.91	2.7	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Silver	<	0.13 mg/kg	U	*	0.13	0.53	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Thallium	<	0.10 mg/kg	U		0.10	0.53	1	11/4/2015	14:30:14	ICPMS2	ICPMS Seq	MS
Vanadium		24.1 mg/kg			0.95	2.7	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS
Zinc		35.5 mg/kg		*	0.22	1.1	1	11/3/2015	01:43:06	ICPMS2	ICPMS Seq	MS

Comments:

-



INITIAL AND CONTINUING CALIBRATION VERIFICATION

Sequence: ICPMS Seq
 Instrument: ICPMS2
 Method: ICPMS Method
 Analyst: BNW

Sample ID	Result ug/L	True Value ug/L	% Recovery	Q	QC Criteria	Analysis Date	Analysis Time	Run Number	CAS Number
ICV001					ICV				
Antimony	186	199	93		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-36-0
Barium	93.8	99.0	95		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-39-3
Beryllium	109	99.0	110		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-41-7
Cadmium	109	99.0	110		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-43-9
Chromium	100	98.0	102		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-47-3
Cobalt	107	100	107		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-48-4
Lead	185	200	93		90 - 110	11/2/2015	20:14	ICPMS Seq	7439-92-1
Manganese	103	100	103		90 - 110	11/2/2015	20:14	ICPMS Seq	7439-96-5
Nickel	105	101	104		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-02-0
Selenium	215	206	104		90 - 110	11/2/2015	20:14	ICPMS Seq	7782-49-2
Silver	109	100	109		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-22-4
Vanadium	95.8	100	96		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-62-2
Zinc	204	205	100		90 - 110	11/2/2015	20:14	ICPMS Seq	7440-66-6
CCV001					CCV				
Antimony	474	500	95		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-36-0
Barium	4720	5000	94		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-39-3
Beryllium	498	500	100		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-41-7
Cadmium	977	1000	98		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-43-9
Chromium	947	1000	95		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-47-3
Cobalt	987	1000	99		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-48-4
Lead	950	1000	95		90 - 110	11/2/2015	20:41	ICPMS Seq	7439-92-1
Manganese	949	1000	95		90 - 110	11/2/2015	20:41	ICPMS Seq	7439-96-5
Nickel	943	1000	94		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-02-0
Selenium	969	1000	97		90 - 110	11/2/2015	20:41	ICPMS Seq	7782-49-2
Silver	151	150	101		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-22-4
Vanadium	948	1000	95		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-62-2
Zinc	993	1000	99		90 - 110	11/2/2015	20:41	ICPMS Seq	7440-66-6



INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Sequence: ICPMS Seq
 Instrument: ICPMS2
 Method: ICPMS Method
 Analyst: BNW

Analyte	Result ug/L	Acceptance Limit	Conc Qual	MDL	RL	Analysis Date	Analysis Time	Run
ICB001								
Antimony	0.28	2.0	J	0.26	2.0	11/2/2015	20:20	ICPMS Seq
Barium	0.043	10.0	U	0.29	10.0	11/2/2015	20:20	ICPMS Seq
Beryllium	0.00	1.0	U	0.30	1.0	11/2/2015	20:20	ICPMS Seq
Cadmium	0.011	1.0	U	0.061	1.0	11/2/2015	20:20	ICPMS Seq
Chromium	0.10	2.0	U	0.24	2.0	11/2/2015	20:20	ICPMS Seq
Cobalt	0.0080	1.0	U	0.026	1.0	11/2/2015	20:20	ICPMS Seq
Lead	0.023	1.0	J	0.015	1.0	11/2/2015	20:20	ICPMS Seq
Manganese	0.006	1.0	U	0.080	1.0	11/2/2015	20:20	ICPMS Seq
Nickel	0.018	1.0	U	0.071	1.0	11/2/2015	20:20	ICPMS Seq
Selenium	0.024	5.0	U	0.88	5.0	11/2/2015	20:20	ICPMS Seq
Silver	0.0030	1.0	U	0.033	1.0	11/2/2015	20:20	ICPMS Seq
Vanadium	0.17	5.0	U	0.35	5.0	11/2/2015	20:20	ICPMS Seq
Zinc	0.050	2.0	U	0.22	2.0	11/2/2015	20:20	ICPMS Seq
CCB001								
Antimony	0.020	2.0	U	0.26	2.0	11/2/2015	20:47	ICPMS Seq
Barium	0.053	10.0	U	0.29	10.0	11/2/2015	20:47	ICPMS Seq
Beryllium	0.00	1.0	U	0.30	1.0	11/2/2015	20:47	ICPMS Seq
Cadmium	0.015	1.0	U	0.061	1.0	11/2/2015	20:47	ICPMS Seq
Chromium	0.14	2.0	U	0.24	2.0	11/2/2015	20:47	ICPMS Seq
Cobalt	0.0070	1.0	U	0.026	1.0	11/2/2015	20:47	ICPMS Seq
Lead	0.014	1.0	U	0.015	1.0	11/2/2015	20:47	ICPMS Seq
Manganese	0.002	1.0	U	0.080	1.0	11/2/2015	20:47	ICPMS Seq
Nickel	0.0090	1.0	U	0.071	1.0	11/2/2015	20:47	ICPMS Seq
Selenium	0.025	5.0	U	0.88	5.0	11/2/2015	20:47	ICPMS Seq
Silver	0.0010	1.0	U	0.033	1.0	11/2/2015	20:47	ICPMS Seq
Vanadium	-0.15	5.0	U	0.35	5.0	11/2/2015	20:47	ICPMS Seq
Zinc	0.039	2.0	U	0.22	2.0	11/2/2015	20:47	ICPMS Seq

PREPARATION BLANK SUMMARY

Sequence: ICPMS Seq
 Instrument: ICPMS2
 Method: ICPMS Method
 Analyst: BNW

Sample ID	Result (mg/kg)	Conc Qual	Q	Acceptance Limit	MDL	Analysis Date	Analysis Time	Run	CAS Number
Method Blank 1				SOIL					
Antimony	0.011	U		0.22	0.22	11/3/2015	01:29	ICPMS Seq	7440-36-0
Barium	0.05	U		1.6	1.6	11/3/2015	01:29	ICPMS Seq	7440-39-3
Beryllium	0.003	U		0.16	0.16	11/3/2015	01:29	ICPMS Seq	7440-41-7
Cadmium	0.0030	U		0.098	0.098	11/3/2015	01:29	ICPMS Seq	7440-43-9
Chromium	0.070	U		0.21	0.21	11/3/2015	01:29	ICPMS Seq	7440-47-3
Cobalt	0.012	U		0.11	0.11	11/3/2015	01:29	ICPMS Seq	7440-48-4
Lead	0.032	U		0.13	0.13	11/3/2015	01:29	ICPMS Seq	7439-92-1
Manganese	0.028	U		0.16	0.16	11/3/2015	01:29	ICPMS Seq	7439-96-5
Nickel	0.009	U		0.12	0.12	11/3/2015	01:29	ICPMS Seq	7440-02-0
Selenium	0.68	U		0.85	0.85	11/3/2015	01:29	ICPMS Seq	7782-49-2
Silver	-0.001	U		0.12	0.12	11/3/2015	01:29	ICPMS Seq	7440-22-4
Vanadium	0.075	U		0.89	0.89	11/3/2015	01:29	ICPMS Seq	7440-62-2
Zinc	0.027	U		0.21	0.21	11/3/2015	01:29	ICPMS Seq	7440-66-6
Method Blank 2				SOIL					
Copper	0.016	U		0.25	0.25	11/3/2015	15:06	ICPMS Seq	7440-50-8
Method Blank 3				SOIL					
Arsenic	-0.010	U		0.19	0.19	11/4/2015	14:14	ICPMS Seq	7440-38-2
Thallium	0.015	U		0.095	0.095	11/4/2015	14:14	ICPMS Seq	7440-28-0



INTERFERENCE CHECK SAMPLE

Client: EISC Laboratories, Inc. SDG No.: Deliverable ID
 Contract: _____ Lab Code: _____ Case No.: _____ SAS No.: _____
 ICS Source: _____ Instrument ID: ICPMS2

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window	Analysis Date	Analysis Time	Run Number
ICSA002								
	Antimony	2.3	1.5	153	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Barium	1.7	1.2	142	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Beryllium	0.31			-2.00 to 2.00	11/3/2015	14:39	ICPMS Seq
	Cadmium	-0.36	0.70	-51	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Chromium	18.1	21.0	86	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Cobalt	1.2	1.0	120	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Copper	9.1	8.0	114	-1,400 - 1,600%	11/3/2015	14:39	ICPMS Seq
	Lead	4.4	4.0	110	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Manganese	7.2	7.0	103	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Nickel	6.8	6.0	113	-100 - 300%	11/3/2015	14:39	ICPMS Seq
	Selenium	2.3	0.30	767	-2,900 - 3,100%	11/3/2015	14:39	ICPMS Seq
	Vanadium	-0.024	0.50	-5	-900 - 1,100%	11/3/2015	14:39	ICPMS Seq
	Zinc	7.0	11.0	64	-100 - 300%	11/3/2015	14:39	ICPMS Seq
ICSAB002								
	Antimony	22.5	22.0	102	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Barium	21.5	22.0	98	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Beryllium	20.7	19.0	109	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Cadmium	19.7	20.0	99	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Chromium	38.0	40.0	95	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Cobalt	21.1	20.0	106	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Copper	28.1	25.0	112	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Lead	23.5	25.0	94	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Manganese	27.6	27.0	102	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Nickel	25.9	24.0	108	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Selenium	21.6	19.0	114	50 - 150%	11/3/2015	14:45	ICPMS Seq
	Vanadium	19.0	19.0	100	80 - 120%	11/3/2015	14:45	ICPMS Seq
	Zinc	23.3	29.0	80	80 - 120%	11/3/2015	14:45	ICPMS Seq
ICSA003								
	Arsenic	-0.29	0.10	-290	-1,400 - 1,600%	11/4/2015	13:43	ICPMS Seq
	Manganese	7.4	7.0	106	-100 - 300%	11/4/2015	13:43	ICPMS Seq
	Thallium	0.12			-2.00 to 2.00	11/4/2015	13:43	ICPMS Seq



MATRIX SPIKE SUMMARY

Sequence: ICPMS Seq

Instrument: ICPMS2

Method: ICPMS Method

Analyst: BNW

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	% RPD	RPD Qual	Storet Codes
Sample 7MS	Client ID: Client Sample 7S				11/3/2015		02:30:02		Matrix:	SOIL		
Antimony	mg/kg	75 - 125	8.1		0.26	U	11.4	71	*			7440-36-0
Arsenic	mg/kg	75 - 125	12.0		6.9		4.6	111				7440-38-2
Barium	mg/kg	75 - 125	308		90.7		229	95				7440-39-3
Beryllium	mg/kg	75 - 125	7.5		0.61		5.7	121				7440-41-7
Cadmium	mg/kg	75 - 125	6.7		0.11	U	5.7	118				7440-43-9
Chromium	mg/kg	75 - 125	42.8		13.3		22.9	129	*			7440-47-3
Cobalt	mg/kg	75 - 125	73.1		15.6		57.2	101				7440-48-4
Copper	mg/kg	75 - 125	37.7		6.3		28.6	110				7440-50-8
Lead	mg/kg	75 - 125	34.0		26.3		2.3	335				7439-92-1
Manganese	mg/kg	75 - 125	720		783		57.2	-110				7439-96-5
Nickel	mg/kg	75 - 125	71.4		9.7		57.2	108				7440-02-0
Selenium	mg/kg	75 - 125	12.0		1.1	J	11.4	96				7782-49-2
Silver	mg/kg	75 - 125	7.2		0.14	U	5.7	126	*			7440-22-4
Thallium	mg/kg	75 - 125	5.3		0.11	J	5.7	91				7440-28-0
Vanadium	mg/kg	75 - 125	82.1		19.7		57.2	109				7440-62-2
Zinc	mg/kg	75 - 125	94.5		27.9		57.2	116				7440-66-6
Sample 9MS	Client ID: Client Sample 9S				11/3/2015		03:23:44		Matrix:	SOIL		
Antimony	mg/kg	75 - 125	7.2		0.39	J	12.6	54	*			7440-36-0
Arsenic	mg/kg	75 - 125	25.2		16.0		5.0	184	*D			7440-38-2
Barium	mg/kg	75 - 125	475		171		252	121				7440-39-3
Beryllium	mg/kg	75 - 125	7.9		0.99		6.3	110				7440-41-7
Cadmium	mg/kg	75 - 125	7.4		0.84	J	6.3	104	D			7440-43-9
Chromium	mg/kg	75 - 125	54.5		26.4		25.2	112	D			7440-47-3
Cobalt	mg/kg	75 - 125	101		31.8		63.1	110	D			7440-48-4
Copper	mg/kg	75 - 125	52.4		24.0		31.5	90				7440-50-8
Lead	mg/kg	75 - 125	38.1		28.7		2.5	376				7439-92-1
Manganese	mg/kg	75 - 125	3930		2840		63.1	1727	D			7439-96-5
Nickel	mg/kg	75 - 125	419		258		63.1	255	D			7440-02-0
Selenium	mg/kg	75 - 125	15.1		1.8	J	12.6	106				7782-49-2
Silver	mg/kg	75 - 125	7.4		0.16	U	6.3	117				7440-22-4
Thallium	mg/kg	75 - 125	5.7		0.13	U	6.3	90				7440-28-0
Vanadium	mg/kg	75 - 125	77.0		20.6		63.1	89	D			7440-62-2
Zinc	mg/kg	75 - 125	736		577		63.1	252				7440-66-6



POST DIGEST SPIKE SUMMARY

SDG No.: Deliverable ID

Sample ID: QJ15014-007

Spiked ID: Sample 7A

Client ID: Client Sample 7A

Matrix: SOIL

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	Analytical Date	Time
Antimony	ug/L	75 - 125	4.4		0.26	U	4.0	110		11/3/2015	03:03:34
Chromium	ug/L	75 - 125	26.5		23.1		4.0	85		11/3/2015	03:03:34
Silver	ug/L	75 - 125	2.3		0.064	J	2.0	112		11/3/2015	03:03:34

- 6 -

DUPLICATE SAMPLE SUMMARY

Client: EISC Laboratories, Inc. Level: LOW SDG No.: Deliverable ID
 Contract: _____ Lab Code: _____ Case No.: _____
 Matrix: SOIL Sample ID: QJ15014-007 Client ID: Client Sample 7D
 Percent Solids for Sample: 86.05 Duplicate ID: Sample 7D Percent Solids for Duplicate: 86.05

Analyte	Units	Acceptance Limit	Sample Result	C	Duplicate Result	C	RPD	Qual	M
Antimony	mg/kg		0.24	U	0.24	U			MS
Barium	mg/kg		90.7		64.3		34	*	MS
Beryllium	mg/kg	0.55	0.61		0.71		15		MS
Cadmium	mg/kg		0.11	U	0.11	U			MS
Chromium	mg/kg		13.3		20.6		43	*	MS
Cobalt	mg/kg		15.6		15.7		1		MS
Copper	mg/kg		6.3		6.0		5		MS
Lead	mg/kg		26.3		26.2		0		MS
Manganese	mg/kg		783		547		35	*	MS
Nickel	mg/kg		9.7		10.5		8		MS
Selenium	mg/kg		1.1	J	1.2	J	9		MS
Silver	mg/kg		0.13	U	0.13	U			MS
Vanadium	mg/kg		19.7		24.0		20		MS
Zinc	mg/kg		27.9		30.1		8		MS

LABORATORY CONTROL SAMPLE SUMMARY

Sequence: ICPMS Seq
 Instrument: ICPMS2
 Method: ICPMS Method
 Analyst: BNW

Sample ID	Units	True Value	Result	C	% Recovery	Q	QC Criteria	CAS Number
LCS 1								
Antimony	mg/kg	2.0	2.1		105		70 - 130	7440-36-0
Barium	mg/kg	10.0	9.4		94		70 - 130	7440-39-3
Beryllium	mg/kg	1.0	1.1		110		70 - 130	7440-41-7
Cadmium	mg/kg	1.0	1.1		110		70 - 130	7440-43-9
Chromium	mg/kg	2.0	2.0		100		70 - 130	7440-47-3
Cobalt	mg/kg	1.0	1.1		110		70 - 130	7440-48-4
Lead	mg/kg	1.0	0.96		96		70 - 130	7439-92-1
Manganese	mg/kg	1.0	1.0		100		70 - 130	7439-96-5
Nickel	mg/kg	1.0	1.0		100		70 - 130	7440-02-0
Selenium	mg/kg	5.0	6.5		130		70 - 130	7782-49-2
Silver	mg/kg	1.0	1.1		110		70 - 130	7440-22-4
Vanadium	mg/kg	5.0	4.8		96		70 - 130	7440-62-2
Zinc	mg/kg	2.0	2.3		115		70 - 130	7440-66-6

LCS 2								
Copper	mg/kg	2.0	2.1		105		70 - 130	7440-50-8

LCS 3								
Arsenic	mg/kg	1.0	1.2		120		70 - 130	7440-38-2
Thallium	mg/kg	1.0	0.94		94		70 - 130	7440-28-0



SERIAL DILUTION SAMPLE SUMMARY

Client: EISC Laboratories, Inc. **SDG No.:** Deliverable ID
Contract: _____ **Lab Code:** _____ **Case No.:** _____ **SAS No.:** _____
Matrix: WATER **Level:** LOW **Client ID:** Client Sample 7L
Sample ID: QJ15014-007 **Serial Dilution ID:** Sample 7L

Analyte	Initial Result ug/L	C	Serial Result ug/L	C	% Difference	Qual	Acceptance Limits	M
Antimony	0.26	U	0.26	U			10.00 %	MS
Barium	158.00		161.00		2		10.00 %	MS
Beryllium	1.10		0.30	U	100.0		10.00 %	MS
Cadmium	0.11	J	0.06	U	100.0		10.00 %	MS
Chromium	23.10		24.80		7		10.00 %	MS
Cobalt	27.30		29.30		7		10.00 %	MS
Copper	10.90		10.90		0		10.00 %	MS
Lead	45.80		45.40		1		10.00 %	MS
Manganese	1360.00		1490.00		10		10.00 %	MS
Nickel	16.90		17.80		5		10.00 %	MS
Selenium	2.00	J	0.88	U	100.0		10.00 %	MS
Silver	0.06	J	0.03	U	100.0		10.00 %	MS
Vanadium	34.30		36.70		7		10.00 %	MS
Zinc	48.60		55.20		14	*	10.00 %	MS



INTERNAL STANDARDS RELATIVE INTENSITY SUMMARY

ICP ID Number: ICPMS2 Run No.: ICPMS Seq Method: MS

Start Time: 1927 End Time: 2056

Sample ID	Client ID	Time	Internal Standards %RI For:											
			Element 209Bi	Q	Element 115In	Q	Element 6Li	Q	Element 45Sc	Q	Element 159Tb	Q	Element 89Y	Q
QJ14026-001	ZZZZZZ	2335												
QJ14026-002	ZZZZZZ	2342												
QJ14026-003	ZZZZZZ	2348												
QJ14026-004	ZZZZZZ	2355												
QJ14026-005	ZZZZZZ	0002												
QJ14026-006	ZZZZZZ	0009												
QJ14026-007	ZZZZZZ	0015												
QJ14026-008	ZZZZZZ	0022												
QJ14026-009	ZZZZZZ	0029												
QJ14026-010	ZZZZZZ	0036												
CCV4	CCV004	0042	77.9		80.6		89.1		83.5		83.3		82.0	
CCB4	CCB004	0049	95.1		95.7		104.0		100.2		94.3		99.7	
QQ88468-001	ZZZZZZ	0056												
QQ88468-002	ZZZZZZ	0102												
QJ29004-001	ZZZZZZ	0109												
CCV5	CCV005	0116	76.0		79.4		89.8		84.3		81.2		82.3	
CCB5	CCB005	0123	94.4		95.7		105.2		103.1		93.6		100.9	
Method Blan	PBS001	0129	94.3		95.3				102.1		93.4		99.9	
LCS 1	LCS	0136	93.9		92.2				99.0		91.4		95.0	
Sample 1	Client Sample 1	0143	93.4		92.9				101.0		91.6		117.4	
Sample 2	Client Sample 2	0149	94.6		93.0				101.2		92.7		116.7	
Sample 3	Client Sample 3	0156	94.1		92.7				100.2		92.7		123.9	
Sample 4	Client Sample 4	0203	94.5		94.0				102.4		95.2		132.9	*
Sample 5	Client Sample 5	0209	95.6		93.7				100.7		94.5		119.1	
Sample 6	Client Sample 6	0216	96.7		95.0				100.2		95.4		112.9	
Sample 7	Client Sample 7	0223	98.7		95.1				101.4		96.3		115.0	
Sample 7MS	Client Sample 7S	0230	98.5		94.4				99.5		95.2		118.0	
CCV6	CCV006	0236	81.0		81.1		84.3		83.2		84.8		82.5	
CCB6	CCB006	0243	98.3		97.7		99.9		101.1		96.8		100.6	
Sample 7D	Client Sample 7D	0250	97.6		93.4				99.9		96.0		115.8	
Sample 7L	Client Sample 7L	0256	101.2		98.1				102.1		97.3		105.3	
Sample 7A	Client Sample 7A	0303	97.3		92.9		96.9		98.7		95.0		111.9	
Sample 8	Client Sample 8	0310	100.5		95.8				99.2		96.9		109.9	
Sample 9	Client Sample 9	0317	96.5		91.4				95.2		95.0		130.5	*
Sample 9MS	Client Sample 9S	0323	96.8		91.7				94.9		95.5		141.0	*
Sample 9DU	Client Sample 9D	0330	97.7		92.1				96.5		96.8		136.8	*
Sample 9L	Client Sample 9L	0337	102.3		95.7				99.4		96.8		106.1	

Comments:

